

NO	ID ALAT	JENIS ALAT	PEMEGANG	DEPT	FACT	MASTER	ALAT	TGL KALIB	EXPIRED	KETERANGAN
1	P-128	THICKNESS	HERI	SYNTHETIC	F	0.5	0.5	04/02/26	01/06/26	OK
2	T-108	THICKNESS	ICHA	GUDANG KULIT	F	0.5	0.5	04/02/26	01/06/26	OK
3	T-31	THICKNESS	RECHA	QC POLA	F	0.5	0.5	04/02/26	01/06/26	OK
4	T-105	THICKNESS		GUDANG KULIT	F	0.5	0.5	04/02/26	01/06/26	OK
5	T-108	THICKNESS	ICHA	GUDANG KULIT	F	0.5	0.51	04/02/26	01/06/26	OK
6	T-103	THICKNESS	IYUS	QC POLA	F	0.5	0.5	04/02/26	01/06/26	OK
7	T-109	THICKNESS			F	0.5	0.51	04/02/26	01/06/26	OK
8	T-104	THICKNESS		QC POLA	F	0.5	0.5	04/02/26	01/06/26	OK
9	P-90	THICKNESS	NANANG	GUDANG KULIT	F	0.5	0.5	04/02/26	01/06/26	OK
10	P-109	THICKNESS	MARDIAN	GUDANG KULIT	F	0.5	0.51	04/02/26	01/06/26	OK
11	P-89	THICKNESS	SUKIRNO	GUDANG KULIT	F	0.5	0.5	04/02/26	01/06/26	OK
12	P-91	THICKNESS	WAHYUDIN	GUDANG KULIT	F	0.5	0.5	04/02/26	01/06/26	OK
13	P-28	THICKNESS	IYUS	GUDANG KULIT	F	0.5	0.5	04/02/26	06/06/26	OK
14	P-129	THICKNESS	DEDE PANDI	GUDANG KULIT	F	0.5	0.5	04/02/26	01/06/26	OK
15	T-73	THICKNESS		QC POLA	F	0.5	0.5	04/02/26	01/06/26	OK
16	P-127	THICKNESS		GUDANG KULIT	F	0.5	0.5	04/02/26	01/06/26	0.5
17	T-102	THICKNESS	RISFANI	GUDANG KULIT	F	0.5	0.5	04/02/26	01/06/26	OK
18	T-102	THICKNESS	RISFANI	GUDANG KULIT	F	0.5	0.5	05/02/26	01/06/26	OK
19	T-73	THICKNESS	969468	GUDANG KULIT	F	0.5	0.5	05/02/26	01/06/26	OK
20	P-85	THICKNESS	LILIA	QC POLA	A	0.5	0.51	05/02/26	01/06/26	OK
21	T-49	THICKNESS		QC POLA	A	0.5	0.5	05/02/26	01/06/26	OK
22	T-33	THICKNESS	LILIS	QC POLA	A	0.5	0.5	05/02/26	01/06/26	OK
23	T-60	THICKNESS	ERUSIN	QC POLA	A	0.5	0.5	05/02/26	01/06/26	OK
24	T-111	THICKNESS		QC POLA	A	0.5	0.5	05/02/26	01/06/26	OK
25	P-120	THICKNESS	AFIT	QC POLA	A	0.5	0.51	05/02/26	01/06/26	OK
26	P-133	THICKNESS	LILIS	QC POLA	A	0.5	0.51	05/02/26	01/06/26	OK
27	P-132	THICKNESS	SANTIANI	QC POLA	A	0.5	0.51	05/02/26	01/06/26	OK
28	P-119	THICKNESS	SANTIANI	QC POLA	A	0.5	0.51	05/02/26	01/06/26	OK
29	T-16	THICKNESS	P Adit		A	0.5	0.5	07/02/26	01/06/26	OK
30	P-134	THICKNESS	P Adit		A	0.5	0.5	07/02/26	01/06/26	OK
31	P-118	THICKNESS	-	SAMPLE	F	0.5	0.5	24/02/26	01/06/26	OK

NO	ID ALAT	JENIS ALAT	PEMEGANG	DEPT	FACT	MASTER	ALAT	TGL KALIB	EXPIRED	KETERANGAN
32	T-59	THICKNESS		SAMPLE	F	0.5	0.5	24/02/26	01/06/26	OK
33	P-122	THICKNESS	SONY	SAMPLE	F	0.5	0.5	24/02/26	01/06/26	OK
34	P-41	THICKNESS		INCOMING	F	0.50	0.51	28/02/26	01/06/26	OK
35	P-106	THICKNESS		INCOMING	F	0.51	0.46	28/02/26	01/06/26	RUSAK
36	P-67	THICKNESS		INCOMING	F	0.5	0.5	28/02/26	01/06/26	OK
37	P-135	THICKNESS	Dadang Bule	PRINTING	B	0.5	0.6	04/02/26	01/06/26	RUSAK
38	P-136	THICKNESS	ANDRIANI	INCOMING	F	0.5	0.5	04/03/26	01/06/26	OK / Baru Pengganti P-106
39	P-137	THICKNESS	Dadang Bule	PRINTING	B	0.5	0.5	17/03/26	01/06/26	OK /Baru Pengganti P-135